

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES

NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above AND Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p>B Flanges: all flanges</p> <p>C Valves: all line valves</p> <p>D Expansion joints, flexible connections, and hose assemblies: all types</p> <p>E Strainers, filters, separators, and steam traps</p> <p>F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p>G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p>H Pressure retaining components that do not fall into one of the above categories</p> <p>N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/>, (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.</p>
SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT	
<p>ROSEMOUNT[®] Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input type="checkbox"/> DESCRIBE OTHER: molded, various</p>
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:	
Model 3300/3400/3500/VP pH sensor product data sheet PERpH-X Sensor ASME Compliance drawing PN 1700749 Proof pressure test report RAI 02 rev 2	

DECLARATION:

I Dana Crowlev (see note 3) employed by Rosemount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowlev*
 Declared before me at Irvine, California, County of Orange
 This 24th day of October AD 2012
 Commissioner of Oaths or Notary Public: (sign) *Catherine Ann Stip*



This space for Regulatory Authority use This registration must be revalidated after ten (10) years from the date of registration.	
CRN: <u>0F12814.2</u> FID#: <u>14714</u>	<p align="center">REGISTRATION ONLY</p> CRN <u>0F12814.27</u> EXAMINER: <u><i>Ken Cross</i></u> for CHIEF BOILER INSPECTOR DATE: <u>02/01/2013</u>
Notes: 1. All fittings shall be registered in the name of the Manufacturer. 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product. 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.	<p><input type="checkbox"/> BLRs <input type="checkbox"/> PVs</p> <p><input checked="" type="checkbox"/> FITTINGS <input type="checkbox"/> COMPONENTS</p> <p align="right">Rev. 1 06/2003</p>

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES


NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above AND Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
<p>CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p>B Flanges: all flanges</p> <p>C Valves: all line valves</p> <p>D Expansion joints, flexible connections, and hose assemblies: all types</p> <p>E Strainers, filters, separators, and steam traps</p> <p><input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p>G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p>H Pressure retaining components that do not fall into one of the above categories</p> <p>N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OR CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.</p>
<p>SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p>ROSEMOUNT[®] Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input type="checkbox"/> DESCRIBE OTHER: molded, various</p>
<p>LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Model 3300/3400/3500/VP pH sensor product data sheet PERpH-X Sensor ASME Compliance drawing PN 1700749 Proof pressure test report RAI 02 rev 2</p>	

DECLARATION:

I, Dana Crawley (see note 3) employed by Rosemount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: Dana Crawley
 Declared before me at Irvine, California, County of Orange
 This 24th day of October AD 2012
 Commissioner of Oaths or Notary Public: (sign) Catherine Ann Stip



This space for Regulatory Authority use

This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: 0F12814.2

FID#: 14714

Date: Feb. 1/13

CRN: 0F12814.28

Dwg: as described

Signed: Yvonne Perry

1 of 1 Part

REGULATION

Fittings Rev. 1 06/2003

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.


UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES


NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above AND Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY.	
<p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p>B Flanges: all flanges</p> <p>C Valves: all line valves</p> <p>D Expansion joints, flexible connections, and hose assemblies: all types</p> <p>E Strainers, filters, separators, and steam traps</p> <p><input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p>G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p>H Pressure retaining components that do not fall into one of the above categories</p> <p>N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/>. (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.</p>
SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT	
	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/></p> <p>CAST <input type="checkbox"/> OTHER <input type="checkbox"/></p> <p>DESCRIBE OTHER: molded, various</p>
LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:	
<p>Model 3300/3400/3600/VP pH sensor product data sheet PERpH-X Sensor ASME Compliance drawing PN 1700749 Proof pressure test report RAI 02 rev 2</p>	

DECLARATION:

I Dana Crowley (see note 3) employed by Rosemount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowley*
 Declared before me at Irvine, California, County of Orange
 This 24th day of October AD 2012
 Commissioner of Oaths or Notary Public: (sign) *Catherine Ann Stip*



This space for Regulatory Authority use	
This registration must be revalidated after ten (10) years from the date of acceptance.	
CRN: <u>OF12814.2</u> FID#: <u>14714</u>	<p>ACCEPTED</p> <p>PROVINCE OF PRINCE EDWARD ISLAND ENVIRONMENT, LABOUR & JUSTICE</p> <p>C.R.N. <u>OF 12814.29</u> DATE: <u>Feb 11 13</u></p> <p align="center">INSPECTION SERVICES SECTION BOILER/PRESSURE VESSEL BRANCH</p>
<p>Notes:</p> <ol style="list-style-type: none"> All fittings shall be registered in the name of the Manufacturer. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation. The declaration shall be made by the person having full authority and responsibility for the quality of the end product. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years. 	

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES


NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: <u>Rosemount Analytical, Inc.</u>	
MANUFACTURERS ADDRESS: <u>2400 Barranca Parkway, Irvine, CA 92606</u>	
PLANT LOCATIONS: <u>Address above AND Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190</u>	
<p>CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/>, (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.</p>
<p>SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p align="center">ROSEMOUNT[®] Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input type="checkbox"/> DESCRIBE OTHER: molded, various</p>
<p>LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Model 3300/3400/3500/VP pH sensor product data sheet PERpH-X Sensor ASME Compliance drawing PN 1700749 Proof pressure test report RAI 02 rev 2</p>	

DECLARATION:

I, Dana Crowley (see note 3) employed by Rosemount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: *Dana Crowley*
 Declared before me at Irvine, California, County of Orange
 This 24th day of October AD 2012
 Commissioner of Oaths or Notary Public: (sign) *Catherine Ann Stip*



This space for Regulatory Authority use
 This registration must be revalidated after ten (10) years from the date of registration.

CRN: 0F12814.2
 FID#: 14714

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

GOVERNMENT OF
 NEWFOUNDLAND
 AND LABRADOR
 Department of
 Government Services

Registered 0F12814.20
 Date 2013 / 2 / 12

GOVERNMENT SERVICES CENTRE
 Engineering and Inspection Services

Registered by *[Signature]*

UNDER THE AUTHORITY OF THE
 PUBLIC SAFETY ACT AND
 THE BOILER, PRESSURE VESSEL AND
 COMPRESSED GAS REGULATIONS

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES

NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.

MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92608

PLANT LOCATIONS: Address above AND Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190

CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY.

A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers
 B Flanges: all flanges
 C Valves: all line valves
 D Expansion joints, flexible connections, and hose assemblies: all types
 E Strainers, filters, separators, and steam traps
 F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters
 G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs
 H Pressure retaining components that do not fall into one of the above categories
 N Nuclear components: Class 1 Class 2 Class 3 . (Meeting CNSC or ASME requirements)

TITLE OF THE STANDARD OF CONSTRUCTION
 ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.

SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT

ROSEMOUNT
Analytical

TYPE OF CONSTRUCTION
 FORGED WELDED WROUGHT
 CAST OTHER
 DESCRIBE OTHER:
 molded, various

LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:
 Model 3300/3400/3500/VP pH sensor product data sheet
 PERPH-X Sensor ASME Compliance drawing PN 1700749
 Proof pressure test report RAI 02 rev 2

DECLARATION:

I, Dana Crowley (see note 3) employed by Rosemount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: Dana Crowley

Declared before me at Irvine, California, County of Orange

This 24th day of October AD 2012

Commissioner of Oaths or Notary Public: (sign) Gatherine Ann Stip



This space for Regulatory Affairs
 This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: 0F12814.2

FID#: 14714

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

NUNAVUT

Boilers and Pressure Vessels Act

REGISTERED

CRN 0F12814.2

Date FEB 01 2013

Signed [Signature]
 Chief Inspector

Territorial Registration Fee

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES


NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosemount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92608	
PLANT LOCATIONS: Address above AND Circuito del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
<p>CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY</p> <p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers B Flanges: all flanges C Valves: all line valves D Expansion joints, flexible connections, and hose assemblies: all types E Strainers, filters, separators, and steam traps <input checked="" type="radio"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs H Pressure retaining components that do not fall into one of the above categories N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/> (Meeting CNSC or ASME requirements)</p>	<p>TITLE OF THE STANDARD OF CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.</p>
<p>SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p align="center">ROSEMOUNT Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/> CAST <input type="checkbox"/> OTHER <input type="checkbox"/> DESCRIBE OTHER: molded, various</p>
<p>LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Model 3300/3400/3500/VP pH sensor product data sheet RERpH-X Sensor ASME Compliance drawing PN 1700749 Proof pressure test report RAI 02 rev 2</p>	

DECLARATION:

I, Dana Crowley (see note 3) employed by Rosemount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: Dana Crowley
 Declared before me at Irvine, California, County of Orange
 This 24th day of October AD 2012
 Commissioner of Oaths or Notary Public: (sign) Catherine Ann Stip



This space for Regulatory Authority use
 This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: OF12814.2
 FID#: 14714

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.

GOVERNMENT OF YUKON

BY [Signature]

CRN OF12814.2

DATE 2/1/13

BOILERS BRANCH

Sect. 1.0 - Fittings Rev. 1 06/2003

UNIFORM STATUTORY DECLARATION FORM FOR THE REGISTRATION OF FITTING DESIGNS

NEW BRUNSWICK
NUNAVUT

NOVA SCOTIA
YUKON

PRINCE EDWARD ISLAND
NORTHWEST TERRITORIES

NEWFOUNDLAND AND LABRADOR

MANUFACTURERS NAME: Rosomount Analytical, Inc.	
MANUFACTURERS ADDRESS: 2400 Barranca Parkway, Irvine, CA 92606	
PLANT LOCATIONS: Address above AND Cirullo del Progreso #27, Parque Industrial Progreso, Mexicali, BC, Mexico CP-21190	
CATEGORY OF FITTINGS TO BE REGISTERED. CIRCLE ONE CATEGORY ONLY	
<p>A Pipe fittings, including couplings, tees, elbows, Ys, plugs, unions, pipe caps, or reducers</p> <p>B Flanges: all flanges</p> <p>C Valves: all line valves</p> <p>D Expansion joints, flexible connections, and hose assemblies: all types</p> <p>E Strainers, filters, separators, and steam traps</p> <p><input checked="" type="checkbox"/> F Measuring devices, including pressure gauges, level gauges, sight glasses, levels, or pressure transmitters</p> <p>G Certified capacity-rated pressure relief devices acceptable as primary over pressure protection on boilers, pressure vessels, piping and fusible plugs</p> <p>H Pressure retaining components that do not fall into one of the above categories</p> <p>N Nuclear components: Class 1 <input type="checkbox"/> Class 2 <input type="checkbox"/> Class 3 <input type="checkbox"/>. (Meeting CNSC or ASME requirements)</p>	<p>CODE OF THE STANDARD OF CONSTRUCTION</p> <p>ASME B31.3 and proprietary standards and product data sheets, supported by proof pressure tests.</p>
<p>SHOW MANUFACTURERS NAME, TRADEMARK, OR LOGO AS IT WILL APPEAR ON THE PRODUCT</p> <p align="center">ROSEMOUNT[®] Analytical</p>	<p>TYPE OF CONSTRUCTION</p> <p>FORGED <input type="checkbox"/> WELDED <input type="checkbox"/> WROUGHT <input type="checkbox"/></p> <p>CAST <input type="checkbox"/> OTHER <input type="checkbox"/></p> <p>DESCRIBE OTHER: molded, various</p>
<p>LIST OF SUPPORTING DOCUMENTATION AND IDENTIFICATION OF THE ACTUAL ITEMS TO BE REGISTERED:</p> <p>Model 3300/3400/3500/VP pH sensor product data sheet PERPH-X Sensor ASME Compliance drawing PN 1700748 Proof pressure test report RAI 02 rev 2</p>	

DECLARATION:

I, Dana Crowley (see note 2) employed by Rosomount Analytical Inc. and being the person having full authority and responsibility for the quality of the end product do solemnly declare that the information contained in this form is true and to the best of my knowledge represents the product for which registration is sought. The dimensions, materials of construction, pressure temperature ratings, and identification markings are in accordance with the herein named standards. I further declare that the manufacture of these fittings is regulated by a Quality Control Program which extends to each plant where fabrication occurs in whole or in part and has been verified by Det Norske Veritas as being suitable for that purpose and I make this solemn declaration conscientiously believing it to be true, and knowing that it is of the same force and effect as if made under oath.

Signature of Declarer: Dana Crowley
 Declared before me at Irvine, California, County of Orange
 This 27th day of October AD 2012
 Commissioner of Oaths or Notary Public: (sign) Catherine Ann Stip




This space for Regulatory Authority use

This registration must be revalidated after ten (10) years from the date of acceptance.

CRN: OF12814.2

FID#: 14714

Notes:
 1. All fittings shall be registered in the name of the Manufacturer.
 2. Each category shall be supported with two Statutory Declaration forms and one copy of supporting documentation.
 3. The declaration shall be made by the person having full authority and responsibility for the quality of the end product.
 4. Quality control programs shall be resubmitted for validation at a maximum interval of five (5) years.



Northwest Territories REGISTERED

UNDER THE AUTHORITY OF THE BOILER AND PRESSURE VESSEL ACT,

C.R.N. OF12814.2T

SIGNED [Signature] Rev. 1 05/2003

DATE Feb 1, 2013

150.00