



Member of the FM Global Group

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmaprovals.com

CERTIFICATE OF COMPLIANCE

ELECTRICAL EQUIPMENT

This certificate is issued for the following equipment:

XS-O2-abcde. X-STREAM O2 Transmitter

Type 4X/IP66

a = Probe Type: 1, 2, 3, 4, 5 or 6

b = Probe Length: 1, 2, 3, 4, 5, 6, 7, 8, 9 or A

c = Mounting Plate: 00, 02, 03, 04, 05, 06, 07, 08, 09, 10, 11, 12, 13 or 99

d = Electronics: 01, 02 or 03

e = Calibration Accessories: 00, 01 or 02

XI-abcdef. X-STREAM Xi Remote Interface

Type 4X

a = Remote Type: 01, 02, 03 or 04

b = Mounting: 00, 01 or 02

c = Cable: 00, 10, 11, 12, 13, 14, 15 or 16

d = Stoichiometric Function: 00, 01 or 02

e = Programmable Reference Function: 00, 01 or 02

f = 850 Deg C Process Function: 00, 01 or 02

Equipment Ratings:

Evaluated for use in unclassified (non-hazardous) locations

FM Approved for:

Emerson Process Management Rosemount Analytical Inc.
Irvine, CA, USA

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

Class 3810	2005
ANSI/NEMA 250	1991
ANSI/IEC 60529	2004

Original Project ID: 3032025

Approval Granted: January 30, 2009

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
091124	December 24, 2009		
RR201535	May 25, 2016		

FM Approvals LLC



J. E. Marquedant
Manager, Electrical Systems

25 May 2016
Date