

November 08, 2012

Attention: Beth Covey
EMERSON PROCESS MANAGEMENT
PO BOX 270658
ST LOUIS, MO

The design submission, tracking number 2012-05373, originally received on August 02, 2012 was surveyed and accepted for registration as follows:

CRN : 0F12814.2 **Accepted on:** November 08, 2012

Reg Type: Addition to Acc. Fitting **Expiry Date:** November 08, 2022

Drawing No. : 1700750 & 1700751 Rev B & A

Fitting type: 389 & 396/398 SERIES PH SENSORS

Design registered in the name of : ROSEMOUNT ANALYTICAL INC

The registration is conditional on your compliance with the following notes:

The pressure and temperature ratings are as indicated by the drawing:

- Models 389 rated 100 psi @ +0/65 C;
- Models 396/VP, R/VP and P/VP rated 150 psi @ +0 / 100 C;
- Models 398/VP, R/VP, and P/VP rated 250 psi @ +0 / 100 C;

This registration is valid only for fittings fabricated at the location(s) covered by the QC certificate attached to the accepted AB-41 Statutory Declaration form. This registration is valid only until the indicated expiry date only if the Manufacturer maintains a valid quality management system approved by an acceptable third-party agency until that date. Should the approval of the quality management system lapse before the expiry date indicated above, this registration shall become void.

An invoice covering survey and registration fees will be forwarded from our Revenue Accounts.

If you have any question don't hesitate to contact me by phone at (780) 433-0281 ext 3347 or fax (780) 437-7787 or e-mail Preda@absa.ca.

Sincerely,



PREDA, DRAGOS, P. Eng.